

# SHOEBOX® PureTest User Guide

*Think Outside of the Booth!®*

# Introduction

Welcome to SHOEBOX® PureTest!

SHOEBOX PureTest is an easy-to-use, automated, ANSI-compliant Type 4 tablet audiometer. It is optimized for hearing health and non-hearing health professionals alike to test outside of a sound booth or mobile clinic.

This user guide will help you learn all the capabilities of SHOEBOX PureTest. If you're looking to get set up quickly, you can view all topics at a glance using our online knowledge base at [help.shoebox.md](http://help.shoebox.md).

If your iPad was provided by SHOEBOX, you'll find the user guides in your iBooks app.

## Prerequisites

Before you start using SHOEBOX PureTest, please check the following requirements:

- Google Chrome with automatic updates (activated by default). **SHOEBOX PureTest is only supported by Google Chrome.**
- Microsoft Windows 7 or later or any other operating system supported by Google Chrome including MacOS 10.13 or later.
- PureTest is only validated for use in full-screen mode and does not support the use of windowed mode introduced in iPadOS 26. Using PureTest in windowed mode may affect the functionality of room scans.
- To ensure optimal working conditions, Standard PureTest Kit Inclusions are listed in [Appendix A](#) of this document.

## Important Safety Instructions



Before you start using SHOEBOX, please read the following safety information:

- There are no known contraindications for the use of the SHOEBOX PureTest medical device software.
- No alarms are used in the system. Error and warning messages are used in the system to inform of an abnormal device state (e.g., no internet connectivity with the server, when attempting to exit the test procedure before save).
- If the test subject experiences any discomfort, the user should stop testing immediately.



SHOEBOX Ltd.  
301-80 Aberdeen St.  
Ottawa, ON, Canada K1S 5R5  
[www.shoebox.md](http://www.shoebox.md)  
1-877-349-9934

SHOEBOX® is a Class II medical device listed with  
FDA (Registration No: 3007307502)  
and Health Canada (License No: 93571).

Intended use: For diagnosis of human hearing loss.

Summary: SHOEBOX® PureTest is an iPad-based  
Type 4 Audiometer (as defined in ANSI/ASAS3.6-  
2018 and IEC 60645-1:2017) intended for  
diagnosis of human hearing loss.



Applicable to headphone and  
external microphone

A serious incident is one that results or could result in death, life threatening illness or injury.  
Please report any serious incident occurring in relation to the use of the SHOEBOX PureTest  
directly to SHOEBOX Ltd. and your local health authority.

IFU-13 18.0 2025/11

Legend:



Catalog Number



Manufacturer



[PureTest IFU](#)



The WARNING label identifies conditions or practices that may present danger to the patient  
and/or user.

A paper copy of this document is available upon request.

# Contents

Introduction	2
Important Safety Instructions	2
Contents	4
Installing PureTest	7
If You Received Your iPad from SHOEBOX	7
If You Provided Your Own iPad	7
Connecting Transducers to a 10 <sup>th</sup> Gen iPad	8
Initial Setup: First PureTest Log In	8
Don't Know Your SHOEBOX Login?	9
Syncing Data	10
Logging Out with Unsynced Data	10
Getting Started with PureTest	13
The Home Screen	13
Notifications	15
Notifications panel	16
Feature updates	16
Using PureTest Offline	17
Establishing your Testing Workflow	18
To access your testing workflow	18
Questionnaire	19
Hearing Test Options	19
Automated Hearing Test	19
Manual Hearing Test	21
Results	22
Test Prep	24
Configuring Your Test Prep Settings in the Portal	24
To Configure Your Test Prep Settings	24


What is biological verification?	25
To Disable Biological Verification	26
Performing A Test Prep Activity	27
Important notes before you begin	27
Step 3: Biological Verification	33
Adding a New Test Subject	42
Creating a Test Subject in the Data Management Portal	42
Adding a Pre-Existing Test Subject to a Project	43
Creating a Test Subject in the PureTest App	43
Completing a Test Workflow	45
Managing Test Results	57
To View Saved Results Data	57
Exporting Results	58
To Export Bulk Results Data	58
Test Prep Reports	59
To Generate a Test Prep Report	59
Setting Device Security Options	61
To Access Your Device Security Options	61
Lock Screen	61
Block Testing with Expired Headphones	62
Configuring Language Settings	63
To Change the Default Language Settings	63
App Settings	65
Hardware	65
To Change or Update Your Headphones:	65
To Change or Update Your Microphone:	66
Removing a Microphone	68
Testing	68
Download Device Settings	68

Testing Offline	69
Workflow	70
About	70
Feature updates	70
Version	71
Terms of Service	71
Privacy Policy	71
Acknowledgements	71
Troubleshooting	72
App is Running Slowly	72
Blocked for Testing Error	73
Checking and Updating Your PureTest App	75
Connecting Transducers to a 10 <sup>th</sup> Gen iPad	76
Crackling Noise from Transducer	78
Reporting a Problem	79
To Report a Problem	79
SHOEBOX Audiometry Specifications	82
Appendix A: Standard PureTest Kit Inclusions	84

# Installing PureTest

## If You Received Your iPad from SHOEBBOX

1. Turn on your iPad.
2. Connect to Wi-Fi from within the iPad settings (**Hint:** Go to **Settings > Wi-Fi**).
3. Tap the PureTest app icon to launch PureTest.



## If You Provided Your Own iPad

1. Turn on your iPad.
2. If this is a brand-new iPad, follow the Apple Installation Wizard to set up.
3. Launch the App Store by tapping the App Store icon.
4. Search for **SHOEBBOX PureTest**.
5. Tap the **GET** button to download and install PureTest.
6. Once installed, the app will appear on your iPad home screen.
7. Tap the PureTest app icon to launch PureTest.


## Is Your iPad Set to the Correct Region?

To confirm your region, go to **Settings > General > Language & Region**.

Done! Now you can proceed with your first PureTest log in.

# Connecting Transducers to a 10<sup>th</sup> Gen iPad


If you're using a 10<sup>th</sup> generation iPad, the first time you plug in a pair of transducers connected by an Apple DAC, a popup will ask you to confirm the type of device.



From this prompt, you must select **Headphones**. If you tap **Other Device** for a pair of transducers, they may not perform as expected. To correct the issue, you will need to follow the steps described under [Troubleshooting](#).

## Initial Setup: First PureTest Log In

1. Tap the PureTest app icon to launch PureTest.
2. Tap the **Get Started** button to set up your software.
3. Log in using your SHOEBBOX email and password, or Single Sign-On credentials for organizations with SSO enabled.
4. If prompted, select which transducer (headphone) to download. Choose the one that matches the barcode on your transducer cord.



- **Note:** If you are using an iPad that does not have a 3.5mm headphone jack, you must use an Apple USB-C to 3.5mm Headphone Jack Adapter to connect the headphones. Currently, PureTest only supports Apple's adapter.
  - **Important:** Do not remove the label on the headphones' cord! You may need your headphones' serial number again later.
5. If prompted, select which external microphone to download. Choose the one that matches the serial number on the base of your microphone.

6. Set your passcode.
7. Enable Microphone Access and Location Access by tapping the **I understand** buttons.
  - **Note:** The microphone is used by PureTest to monitor background noise during testing, and location monitoring is used to help organize your data. To run PureTest optimally, please enable both.

### **Don't Know Your SHOEBOX Login?**


**Hint:** The email is the one where you receive your SHOEBOX emails, and the password is the one you created on the SHOEBOX web portal. If you don't know what these are, contact [support@shoebox.md](mailto:support@shoebox.md).

If you have forgotten your password, click the **Forgot Password** link on the login screen and enter the email associated with your SHOEBOX account. You will receive an email with directions on how to reset your password. **Remember that this will also change your password for the SHOEBOX web portal!**

# Syncing Data

SHOEBOX PureTest automatically syncs test results to your Data Management Portal when connected to Wi-Fi. However, if you log out with unsynced data or conduct tests offline, you will need to reconnect to the internet to allow the app to sync the data automatically.

The images below show a comparison of PureTest with all data synced and with unsynced data. The red indicator refers to the number of items still waiting to be synced. This includes test prep activities, and the steps completed during a test workflow.






For more information on offline testing, see [Using PureTest Offline](#).

## Logging Out with Unsynced Data

To allow for greater flexibility and efficiency in managing multiple users, you can log out of PureTest at any time, even if there is unsynced data, and allow another user to log in and continue testing.

A warning will appear to alert you if another user has any unsynced data. Any unsynced data will remain on the device until you log back in and continue the syncing process. This ensures that your data is securely stored and ready for synchronization, maintaining the integrity and completeness of the data.



If you attempt to log out with unsynced items, you will receive a warning message letting you know that your data will never be uploaded unless you log in again when you have regained internet access.

If you do choose to log out with unsynced data, you will need to regain internet access and log in again for that data to be synced.

**Note:** You cannot switch users or log in when offline. You must be connected to Wi-Fi.

**Important:** If, for any reason, the app needs to be uninstalled or deleted, each individual user must log in to sync their data to prevent data loss. Deleting or uninstalling the app will result in the loss of any unsynced data.

## What Happens If a User with Un-Synced Data Was Disabled While They Were Logged Out of the App?


In this situation, admin users will need to follow the steps listed below:

1. Log in to the [SHOEBOX Data Management Portal](#).
2. In the side menu, hover over the **Settings** tab.
4. Under the **Administration** heading, click **Users**.
5. Find and click the name of the user that was disabled.
5. In the top-right corner of the user profile, click the **Enable** button.
6. Click the **Reset Password** button and proceed with resetting the user's password.
7. Once the user has been enabled and their password has been reset, their credentials must be used to log in to the app using the same iPad they logged out from. Once they have logged in with a stable internet connection, their unsynced data will be synced, resolving the issue.
8. The user can then be disabled again in the SHOEBOX Data Management Portal by an admin.

# Getting Started with PureTest

## The Home Screen

After you finish setting up your software, you'll find yourself on the home screen.




**Access Settings Screen:** Tapping here will take you to the app's **Settings** screen, where you can view, configure, and update your hardware, view and download your offline project data, and access other app info such as your app version and terms of service.


**Access Test Prep Screen:** Tapping here will take you to the **Test Prep** screen. This is where you will go to perform the regular pre-test room scans and headphone checks required before beginning your testing session.

**Add Someone New:** Tapping on the **Plus** icon  will open a panel, where you can add a new subject by entering their demographic information.

**Internet, Cloud Sync, Offline Testing Status Indicators:** Informs you about your current internet connection status, your last portal sync, and if there are offline projects available.




**Test Subject Search:** With PureTest, long and cumbersome test subject lists are a thing of the past. Now, you can easily find the individual you're looking for by typing their name or employee ID into the search bar.




Searching for a subject will allow you to view their last tested date (which is updated after every workflow completion, online or offline), and tapping on the subject will show you which project they are in, plus their full demographic information, which you can edit by tapping the **Edit** button.

**Note:** Audiogram history can only be viewed on a patient's page in your web portal.




**Account Panel:** Tapping here will show the name and email address of the user currently logged into the app, and this is also where you will log out and switch users, or select the default language for the app. Additionally, it contains a link to the SHOEBOS Help Center website if you require assistance, and the Report a Problem feature for easy troubleshooting.



# Notifications

Notifications in PureTest are displayed in a carousel on the home screen and ensure you are kept informed about important updates and critical warnings.



You can scroll through multiple notifications by swiping left or right and dismiss a notification by clicking the **X** in the corner of the message. Dismissed messages will be stored in the **Notifications** panel of PureTest for future reference.


When you have new notifications, the **Bell** icon at the top right of the screen will display an exclamation mark. Tapping the **Bell** icon will open the **Notifications** panel.



# Notifications panel


The **Notifications** panel has two tabs:

- **General:** Stores all relevant errors, warnings, and notices for PureTest.
- **Feature Updates:** Stores updates about new features in PureTest.



# Feature updates

When viewing a feature update, clicking **Learn More** will take you to a screen that provides more in-depth information on the feature and instructions on how to use it.



You can also view the latest feature updates from the **About** section on the **Settings** screen. For more information, see [Feature updates](#).


# Using PureTest Offline


PureTest allows users to conduct tests, generate results reports, and perform shift analyses while offline. Offline testing is **enabled by default on newly created projects**.

To enable offline testing for pre-existing projects, you need to use your SHOEBOX web portal to select which of your projects you would like to make available for offline use. While using PureTest offline, you can only perform test functions on subjects that have been placed into offline-enabled projects.

**Note:** Only admin users can create projects and control which projects can be made available for offline testing.

Admin users can enable offline testing by proceeding with the following steps:

1. Log in to your [SHOEBOX Data Management Portal](#).
2. Click the **Projects** tab.
3. Click on the name of the project you want to make available offline.
4. Click the **Edit** icon .
5. Enable the **Make patients available for PureTest offline testing** toggle switch.
6. Click **Save**.



The screenshot shows the 'Projects' section of the SHOEBOX Data Management Portal. The main content area displays the 'Project Details' form for a project named 'project name Active'. The form includes the following fields and controls:

- Name \***: A text input field containing 'project name Active'.
- Description**: A text area for project description.
- Start Date**: A date picker set to 2020-08-24.
- End Date**: A date picker set to 2020-09-30.
- Make patients available for PureTest offline testing**: A toggle switch that is currently turned on (green).
- Status**: A dropdown menu set to 'Open'.

Buttons for 'Cancel' and 'Save' are visible at the top right of the form. A yellow box highlights the toggle switch.

**Important:** If you are planning on testing offline, ensure that you have synced your offline projects by going to **Settings > Testing Offline** and tapping on **Download** while your iPad is online (a strong internet connection is recommended, weak connections will cause the app to switch from online to offline mode sporadically).


# Establishing your Testing Workflow

Before getting started with your testing, you will need to access your SHOEBOX Data Management Portal to configure your testing workflow.


**Note:** Initial, default options are pre-set by SHOEBOX based on your organization's industry. Additionally, changing these settings will apply changes to **all** devices in your organization.

## To access your testing workflow

1. Log in to your [SHOEBOX Data Management Portal](#).
2. Hover over the **Settings** tab in the sidebar.
3. Click on **Testing Workflow**.



On this page you can see a list of all testing workflows configured for your organization. From here you can manage all your workflows by editing, deleting, duplicating, and/or enabling/disabling them. To edit an existing workflow, simply click on its name.




The settings on this page configure the activities that subjects will experience during their testing session.

The settings are displayed in order from start to finish, indicated by the **start of workflow** and **end of workflow** markers at the top and bottom of the page.

By default, the full PureTest workflow is **Questionnaire > Pure Tone Test > Results**. However, both the questionnaire and results can be disabled if they aren't a good fit for your organization's current testing workflow.

## Questionnaire

This section allows you to either select a pre-test questionnaire that you would like your subjects to complete or disable the questionnaire and start the testing workflow with the automated hearing test.



The screenshot shows a settings panel for the 'Questionnaire' section. At the top right, there is a green toggle switch with a checkmark, indicating it is turned ON. Below the toggle, there is a label 'Questionnaire type' followed by a dropdown menu. The dropdown menu is open, showing 'Hearing Conservation 3' as the selected option.

**To disable the questionnaire**, simply leave the toggle switch OFF.

**To enable a questionnaire**, turn the toggle switch ON and select a questionnaire from the dropdown menu.


## Hearing Test Options

Here you can select between the standard automated hearing test and the advanced manual testing option. Each of these test types will have different configuration options.

**Note:** The manual test should only be administered by certified test examiners or audiologists.

### Automated Hearing Test


**Frequencies (Hz):** Select the frequencies you would like to use for the test. The default recommended settings are already configured for you.



The screenshot shows a grid of radio buttons for selecting frequencies. The frequencies are arranged in three columns. The first column contains 250, 500, 750, 1k, and 1.5k. The second column contains 2k, 3k, 4k, 6k, and 8k. The third column contains 10k, 12.5k, and 16k. The radio buttons for 500, 1k, 2k, 3k, 4k, 6k, and 8k are checked.


**Note:** Frequencies are presented from the lowest (250 Hz) to the highest (16k Hz).

**First Frequency Retest:** This feature will retest the first frequency of the hearing test (1k by default). This allows the test subject to get familiar with how the test works, and how to appropriately respond, allowing any mistakes they make on the first frequency to be retested for accuracy. This will reduce the number of incomplete tests and the need to retest any test subjects with erroneous results. The frequency that was retested will be included in the Audiogram Export CSV.



This feature is enabled by default but can be disabled by clicking the toggle switch (pictured above). However, it is recommended to leave this feature enabled. Additionally, the first frequency retest is a requirement for OSHA compliance. If your organization is performing OSHA testing, this feature must remain enabled.

**Test theme:** Select how the test will be displayed to the subject.



**Instructions:** Allows you to enable or disable the pre-test instructions for test subjects. If enabled, the workflow will provide brief instructions for the test subject on how to complete the automated test. If disabled, these instructions will not appear in the workflow.

**Volume:** Allows you to set the minimum and maximum volume of the tones presented in the automated test. The default recommended settings of 0 db HL minimum and 90 db HL maximum are already configured for you (for occupational hearing testing).

**Noise alert mode:** Controls whether you're notified when noise levels are high enough to potentially interfere with test results.

- A. **Notify Immediately** will interrupt the test to notify you and will include ambient noise notation on your audiogram. This is the default setting.
- B. **Delay Notification** will notify you at the end of the test. Select this option if you want to see ambient noise notation on your audiogram, but do not want the test to be interrupted.
- C. **Don't Notify** will not notify you at all and will not print the noise alerts in the interpretation section of the test results.

**Main tone type:** Allows you to choose what type of tone is presented to subjects as the main test tone. You can choose between a pure tone (default), warble tone, or pulsed tone.

**Auto-retest tone:** Allows you to choose what type of tone is presented to subjects as the retest tone. Auto retest tones are presented if the test detects that a subject needs to be retested at a certain frequency. You can choose between a pure tone, warble tone, or pulsed tone (default).

Instructions  Enabled

Volume  Minimum  Maximum

Noise alert mode

Main tone type

Auto-retest tone

## Manual Hearing Test

**Note:** A manual test is only available if it has been configured in the portal.

**Hearing Test Settings**

**Automated Hearing Test**  Patient-driven hearing test, requiring minimal intervention from the examiner.

**Manual Hearing Test**  Trained audiologist/examiner-driven hearing test, requiring manual adjustments to record patient results.

**Frequencies Selection**  
Select the frequencies you want available for testing, and select the ones you want to recommend your examiners to test.

Frequencies (Hz)	Available Frequencies (Hz) ?	Recommended Frequencies (Hz) ?
	<input checked="" type="checkbox"/> 250	<input type="checkbox"/> 250
	<input checked="" type="checkbox"/> 500	<input type="checkbox"/> 500
	<input checked="" type="checkbox"/> 750	<input type="checkbox"/> 750
	<input checked="" type="checkbox"/> 1k	<input type="checkbox"/> 1k
	<input checked="" type="checkbox"/> 1.5k	<input type="checkbox"/> 1.5k
	<input checked="" type="checkbox"/> 2k	<input type="checkbox"/> 2k

**Frequencies (Hz):** The frequencies which will be available for testing during the manual test. You can configure both the available frequencies as well as specific recommended frequencies. A recommended frequency will prompt a reminder to the examiner if it is not tested.

**Main tone type:** Allows you to choose what type of tone is presented to subjects as the main test tone. You can choose between a pure tone (default), warble tone, or pulsed tone. The examiner can change this value during the exam.

**Volume steps:** This setting controls the interval that is changed each time the examiner plays the increase/decrease volume buttons on the test interface.

**Volume:** Allows you to set the minimum and maximum volume of the tones the examiner will be able to present. The default recommended settings of 0 db HL minimum and 90 db HL maximum are already configured for you (for occupational hearing testing).

**Default Tone Presentation Level:** The starting volume when an examiner begins a new test. Typically, the value you want to test first.

The screenshot shows a configuration window with the following sections:

- Tone Type**  
Set the tone type that the test will launch with. Examiners will still have access to other tone types on the testing screen.  
Default tone type: Pure tone (dropdown menu)
- Volume Steps**  
Configure the steps at which examiners can increase or decrease the volume.  
Volume up step:  +5 dB HL,  +10 dB HL  
Volume down step:  -5 dB HL,  -10 dB HL
- Volume**  
Control the range of volumes that the examiner is able to present, and set a default tone presentation level (volume) when launching a new manual test.  
Volume: Minimum (-10 dB HL dropdown), Maximum (90 dB HL dropdown)  
Default tone presentation level: 50 dB HL (dropdown menu)

## Results

This section controls how the test results will be presented at the end of your workflow. You can configure what will be displayed in the test results by turning the following features on or off.

**Primary Identifiers:** Allows you to check or uncheck which subject demographics you would like to be displayed in the results report.

**Shift analysis:** Analyzes whether a shift is present in the employee's hearing. This feature is available online or offline. Note that the shift analysis will only be displayed if a baseline is

present. Baselines are set on a subject's individual patient page in the data management web portal.

**Room scan information:** Includes room scan information in the test results.

**Established baseline:** Displays the employee's most recent baseline if one is available.

**Hearing grade overlay:** Colored shades on the audiogram indicating degrees of hearing loss.

**Explanation notes:** Provides the subject with a brief explanation of whether their hearing changed, and what their next steps should be. Each of these notes can be enabled or disabled as you prefer.

**Signatures:** This section allows you to enable or disable confirmation signatures at the end of the testing workflow. You can enable/disable a confirmation signature for the examiner, the subject, or both, depending on your organization's needs. You can also customize the disclaimer presented before the signature prompt using the text fields. You can use a custom disclaimer in any of the languages supported by PureTest, but it is recommended to use reviewed translations only.

**Results** ON

**Primary identifiers**

- Date of Birth
- Employee ID
- Gender

**Shift analysis** ON Enabled

**Room scan information** ON Enabled

**Established baseline** ON Enabled

**Hearing grade overlay** ON Enabled

**Explanation notes** ON Enabled

- Hearing change explanation**  
E.g. Your hearing test suggests a potential 'Standard' Threshold Shift in your right ear - a change in your hearing when compared to your baseline test.
- Next Steps**  
E.g. You may be asked to do another hearing test to confirm this change.
- Safety statement**  
E.g. Remember to wear your hearing protection when you're in test settings.

We recommend reviewing the translations of the Examiner and Subject Signature text.

**Examiner signature** ON Enabled

**English**

By signing, I confirm the test is complete and that I have shared the initial results with the test subject. I have informed the test subject of any next steps.

300 characters max

**French**

En signant, je confirme que le test est terminé et que j'ai partagé les résultats initiaux avec le sujet du test. J'ai informé le sujet du test de toute étape ultérieure.

300 characters max

**Spanish**

Al firmar, confirmo que la prueba está completa y que he compartido los resultados iniciales con el sujeto de la prueba. Le he informado al sujeto de la prueba sobre los siguientes pasos.

300 characters max

**Subject signature** ON Enabled

**English**

By signing, I agree that I have seen my initial hearing test results. These results can change after professional review. If needed, I will be informed of any next steps.

300 characters max

**French**

En signant, j'accepte que j'ai consulté les résultats initiaux de mon test d'audition. Ces résultats peuvent changer après un examen professionnel. Je serai informé(e) de toute étape ultérieure, le cas échéant.

300 characters max

**Spanish**

Al firmar, confirmo que he visto los resultados iniciales de la prueba de audición. Estos resultados pueden cambiar después de la revisión por parte de un profesional. De ser necesario, me informarán sobre los siguientes pasos.

300 characters max

# Test Prep

The test prep screen will allow you to perform routine hardware checks to **ensure your device is functioning properly**, and in some cases, to maintain compliance with your regulatory requirements.

## Configuring Your Test Prep Settings in the Portal

We recommend you first configure and/or confirm your test prep settings in your SHOEBOX Data Management Portal. The settings you configure in the portal should reflect your organization's regulatory requirements for routine hardware checks. These can be configured to be required on a regular interval. If they are, you will be required to pass the activities prior to testing.

**Important:** Initial, default options are pre-set by SHOEBOX based on your organization's industry. Additionally, changing these settings will apply changes to all devices in your organization. Only make changes to these settings if you are sure of your organization's required configurations.

### To Configure Your Test Prep Settings

1. Log in to your [SHOEBOX Data Management web portal](#).
2. Hover over the **Settings** tab in the sidebar.
3. Under the **PureTest** heading, select **Test Prep**. On this page, you will find the configurations for your ambient noise room scan, your headphones check, and your biological verification.

The screenshot shows the 'Test Prep' settings page in the SHOEBOX Data Management Portal. The page is titled 'Settings | PureTest' and 'Test Prep'. It features a sidebar on the left with navigation options: DASHBOARD, PATIENTS, PARTICIPANTS, REPORTS, DATA ADMIN, FILTERS, PROJECTS, and SETTINGS (highlighted). The main content area is divided into three sections: Room Scan, Headphones Check, and Biological Verification. Each section has a title, a brief description, and several configuration options. The 'Room Scan' section includes 'MPANL Protocol' (with radio buttons for OSHA and ANSI), 'Successful room scan interval' (set to Daily), and 'Require external microphone' (checked). The 'Headphones Check' section includes 'Successful headphones check interval' (set to Daily). The 'Biological Verification' section includes a toggle for 'Biological Verification' (checked), 'Frequencies (Hz)' (with checkboxes for 250, 500, 1k, 2k, 3k, 4k, 6k, and 8k), 'Successful biological verification interval' (set to Daily), and 'Noise alert mode' (set to Notify Immediately). There are 'CANCEL' and 'SAVE' buttons in the top right corner.

- In the **Room Scan** section, first select the **MPANL** (Maximum Permissible Ambient Noise Level) setting that applies to your organization, then use the dropdown menu to select the **interval** at which your app will require a successful room scan. Finally, use the toggle switch to select whether an external microphone is required for noise monitoring.

**Note:** To maintain compliance for occupational hearing testing, select **OSHA** as your MPANL protocol, **Daily** as your successful room scan interval, and set the **Require External Microphone** switch to ON.

**Room Scan**  
*An ambient noise scan of the testing area. This ensures that your environment is quiet enough for testing.*

MPANL Protocol  OSHA  ANSI

Successful room scan interval

Require external microphone

- In the **Headphones Check** section, use the dropdown menu to select the **interval** at which your app will require a routine check.

**Note:** For occupational hearing testing, you would select Daily to maintain compliance.

**Headphones Check**  
*An activity to test that all parts of your headphones are working well.*


Successful headphones check interval

- In the **Biological Verification** section, select the **frequencies** presented during the biological verification, then select the **interval** at which a biological verification is required, and finally the noise alert mode during the biological verification.

**Note:** Changing the test prep settings in the portal for bioverification by adding or removing frequencies will clear shared baselines. A warning will appear to notify the admin user that this will take place. This is to ensure that shared baselines are always consistent with the test prep settings.

## What is biological verification?

Biological verification is a hearing test the examiner performs on themselves to make sure the audiometer is functioning properly. Future results are compared to the first test to confirm that there have been no significant changes. This activity works best when the examiner has stable hearing.




**Note:** The frequencies and noise alert settings configured here should be the same settings your organization uses for hearing testing.


## To Disable Biological Verification

If you are certain that your organization does not need to perform biological verification, you can disable biological verification entirely by clicking on the green toggle switch to turn it off. This will remove biological verification from the test prep page.

A warning notification will let the user know that biological verification is required for compliance in many hearing conservation programs.




If biological verification is toggled off, the test prep screen will look like the image below, containing only the room scan and headphones check activities



# Performing A Test Prep Activity

## Important notes before you begin

- Before you begin your testing for the day, make sure you are logged in under your own user account.
- Ensure you select a room that can serve as a suitable testing environment. The room should be free of any background noise or potential disturbances and should contain a table and chairs for the subject and examiner to use.
- The iPad should be placed in the provided iPad stand (alternatively, if your case has a stand, you can use that) on a secure table. It should **not** be held in one's hands or placed flat on the table as this can cause vibrations that will trigger the noise monitoring. **Ensure the table and iPad are not located near any vents, fans, or appliances like refrigerators that emit noise.**
- If you receive this message telling you to add new headphones, it means your old headphone data has been removed after the headphones were sent for recalibration. You need to install new headphones to the app to start the Test Prep Activities. Contact an admin user if you do not have the new headphones.




In the PureTest app, tap **Test Prep** in the bottom left corner to access the **Test Prep** screen.


If you have configured your test prep steps to be required, you will be blocked from testing until they have been passed successfully. If you have them set to **Never**, you can still perform the test prep activities but will not be blocked from testing if you do not perform them.

For occupational hearing testing, we recommend following the default setting of **Daily**. This will require you to perform a test prep every 24 hours.

**Note:** If your test prep settings are configured to be required at a regular interval, **you must complete the test prep activities in order from Step 1 to Step 3.** If you have them set to **Never**, then the order does not matter.



**Note on biological verification:** Some organizations are not required to perform biological verification as part of their test prep, and the biological verification step may be removed by a system administrator. If you are not required to perform biological verification, your test prep screen will look like the image below and you will only be required to perform steps 1-2.




## Step 1: Room Scan

Before proceeding, make sure the iPad is placed in the iPad stand provided, with the microphone unobstructed. This will ensure that the noise monitoring functions correctly. The iPad should not be placed flat on the table or held in one's hands as this can cause vibrations that will trigger the noise monitoring.


**Note:** PureTest is only validated for use in full-screen mode and does not support the use of windowed mode introduced in iPadOS 26. Using PureTest in windowed mode may affect the functionality of room scans.

### To perform a room scan:


1. Tap the **Room Scan** button. The button will also show you the date of your last room scan on the right.
2. If you have an external microphone, you will then be prompted to plug it in. Once it is plugged in, tap **Next**.



3. The PureTest app will then perform the ambient noise room scan. Once the scan is complete, the app will let you know if the scanned area is a suitable testing environment based on the level of ambient noise detected. If the room scan detects any ambient noise above the maximum levels, the app will indicate which frequencies failed the room scan and will ask for a re-scan.
  - **Note:** The test will also indicate if one of your frequencies passed but was close to failing.



- After confirming the results, tap **Done**. You will then be prompted to remove your external microphone (see below image).




- Once you have passed a room scan, the **Headphones Check** button will be unlocked.

**Important:** Subject hearing tests **must be performed in the same room where you performed this room scan**. Proceeding with testing in an unsuitable environment will produce invalid results.


## Step 2: Headphones Check

**Note:** Ensure that the examiner performing the headphone check has not undergone a change in hearing ability that may compromise their ability to perform a proper headphone check. If necessary, be prepared to double-check with another test examiner.


1. Tap the **Headphones Check** button to begin your headphones check. You will be prompted to plug in your headphones.




**Warning:** Admin users can block testing with expired headphones across the entire organization. This is done to maintain regulatory compliance and prevent misdiagnosis of test subjects. If your organization is blocked from using expired headphones, you will not be able to perform the headphones check, nor will you be able to perform any subject hearing tests. Install new headphones to complete the verification. Contact the admin user if you do not have the new headphones.



- The headphone check will begin with the left speaker. Tap on the **Play** button and the app will present a tone.
  - If the tone was loud and clear, tap the **Pass** button.
  - If the tone presents with any crackling, static, or intermittent sounds, tap the **Fail** button.
- Once you have confirmed the left speaker, repeat the process for the right speaker.



- Once you have confirmed that both the left and right speakers are functioning correctly, tap the **Done** button on the results panel. This will complete your daily headphone check.



**Note:** If during your testing session, you notice that the headphones have received, or may have received, some form of damage, you can always return to the test prep screen and redo the headphones check. This is good practice to ensure that your headphones are always functioning correctly.

- Once you have passed a room scan, the **Biological Verification** button will be unlocked (if biological verification is enabled.) and you can proceed to Step 3. If biological verification is disabled, it will not be present on the test prep screen, and your test prep session will now be complete.

### Step 3: Biological Verification


Biological verification is a hearing test the examiner performs on themselves to make sure the audiometer is functioning properly. Future results are compared to your first test to confirm that there have been no significant changes.

This activity works best when the examiner has stable hearing. If there is more than a 10dB difference at any frequency, a full acoustic calibration may be required.

#### Important:


- If you will be testing offline, ensure the examiner that will be performing the biological verification is logged into the app before going offline. If the examiner is in a situation where they need to perform a biological verification offline, but the app is logged in under another examiner's profile, one solution is to create a hotspot using a mobile device with data access and connecting through that to switch accounts. For more information on how to do this, check out [this quick, simple guide from cnet.com](#) on how to tether using a personal hotspot on an iPhone or Android.
- If the test examiner whose baseline is set on the app is unavailable to perform the daily biological verification (sick, took the day off, etc.) another test examiner may log in to the PureTest app and perform their own biological verification instead. Baselines for each examiner are recorded and used separately.

**Warning:** Admin users can block testing with expired headphones across the entire organization. This is done to maintain regulatory compliance and prevent misdiagnosis of test subjects. If your organization is blocked from using expired headphones, you will not be able to perform the biological verification. Install new headphones to complete the verification. Contact an admin user if you do not have the new headphones.




## Performing your first Biological Verification and Setting a Baseline

1. Tap the **Biological Verification** button to begin your biological verification.




2. You will first be asked to confirm that you are logged in under the correct account to prevent erroneous results. Tap **Yes, continue** if your name appears onscreen.




3. You will then be prompted to allow PureTest to store your baseline in the secure SHOEBOS cloud. This is done so that you can perform rapid tests for your future biological verifications.

Tap **Allow** to proceed while allowing PureTest to save your baseline to the secure SHOEBOS portal.


If you select **Don't Allow**, you will need to set your baseline every time you perform the test and will not be able to use rapid testing.



4. You will then be asked to verify whether the correct headphones are plugged in. Compare the serial number displayed on screen to the serial number on your headphones. If they match, tap **Start Check** to proceed. If they are different, tap **My Serial is Different** to troubleshoot the issue




5. The app will then prompt you to put on the headphones, and to place the iPad in the provided iPad stand. Once that's done, tap **Start the Test!** to begin.



- The test screen will then appear, and you will be prompted to complete it. This is done to ensure that all tones are presenting correctly at all frequencies.

**Note:** Biological verification uses Assisted Mode instead of Classic Mode as the test theme. A test examiner performing biological verification is confirming whether they can hear a sound, and viewing the presentation level on screen **will not impact** the usefulness of the exercise.



Using Assisted Mode is similar to how audiologists perform biological verification in a clinical setting and is not cheating the test. They are not blinded to the test and are also able to see the presentation level when testing their own hearing.



See our [customer help center](#) for more information on completing your first bio verification using assisted mode.

Upon completion, you will be prompted to remove the headphones and will be shown the results from your biological verification. These results will serve as your baseline, and future biological verification results will be compared against this baseline. Until you set your first baseline, you will be blocked from testing patients if biological verification is required.

Tap **Set Baseline** to confirm your baseline, then tap **Done** to finish your first biological verification. You will then be able to proceed with your testing workflows for the day.




**Important:** Contact an admin user if you have had a change in hearing and cannot complete future biological verification checks. An admin user can reset your baseline in the portal, which will force you to set a new baseline in the app the next time you perform a biological verification.

**Note:** Biological verification results are displayed in **tabular format only**. Biological verification results are saved in the app's cache storage until synced with the web portal and are currently **not viewable in the portal**. The baseline value is tied to your account and stored as a part of your account data, unless it is reset by an admin user.


## Performing Regular Biological Verifications using the Rapid Test Feature

After setting a baseline, your future biological verifications are completed using a rapid test for efficiency.


1. Tap **Biological Verification** to begin.




- You will be asked whether you are logged in under your account and to confirm when your baseline was set. If the information on screen is accurate, tap **Yes, continue**.





- Your baseline will be displayed on screen. Tap **Start Check** if it looks correct.



- You will then be asked to verify if the correct headphones are plugged in. Compare the serial number displayed on screen to the serial number on your headphones. If they match, tap **Start Check** to proceed. If they are different, tap **My Serial is Different** to troubleshoot the issue.



- Instructions for the rapid test will be displayed on screen. Tap **Next** and **Done** once you have read and understood the instructions.





- The app will then prompt you to put on the headphones, and to place the iPad in the provided iPad stand. Once that's done, tap **Start the Test!** to begin.



Please **put the headphones on** and for improved accuracy of results **place the iPad on a stand**.



7. The rapid test will begin, starting with the right ear. Follow the on-screen instructions to complete the test.



**Note:** if you respond with **No** to any frequency, you will be given a warning message, and an option to retry the frequency or end the test. If you end the test, you will need to start from the beginning. If you retry the frequency, and you answer **No** again, the biological verification will fail, and you will be given some troubleshooting tips. You will always have the option to recheck.



8. Once you have passed every frequency, you will be taken to the results screen. Tap **Done** to complete the biological verification. You will then be able to proceed with your testing workflows for the day.


# Adding a New Test Subject

New subjects can be added in one of two ways:

1. Created on your SHOEBOX Data Management Portal and imported into your app.
2. Created in the PureTest app.

## Creating a Test Subject in the Data Management Portal

1. Log in to your [SHOEBOX Data Management Portal](#).
2. Hover over **Patients** in the sidebar and click **All Patients**.
3. Click **+New**.



4. Add the relevant demographics. Required fields (first name, last name, project) are marked with an asterisk \*.
  - **Note:** If doing shift calculations, entering the subject's birth date and gender are required.

5. Click **Create** to create the new test subject.

**Note for Admin Users:** If you use Projects, ensure you add your new subject to a project on the portal as this cannot be edited in the app.

## Adding a Pre-Existing Test Subject to a Project

1. Click **Projects** in the sidebar to access the **Projects** page.
2. Click on the **name of the project** you would like to add your subject(s) to.
3. Click **+Add**.


The screenshot shows the 'Example Project' page. At the top right is an 'EXPORT' button. Below is a 'Project Details' section with fields for Start date (Oct 16, 2020), End date (Oct 16, 2028), Status (Open), and Available for PureTest offline testing (Yes). Below this is an 'Import Patients and Audiograms' section with instructions to drop a CSV file and a link to 'Change Date Format'. At the bottom is a table with two tabs: 'Patients (0)' and 'Users (1)'. The 'Patients (0)' tab is active, showing a table with columns for Patient Name, Documents, and Last Tested. A search bar is at the top right of the table. A '+ Add' button with a gear icon is highlighted with a yellow box. Below the table, it says 'There were no results matching your filter criteria.'


4. Select the subject(s) you would like to add to the project.
5. Click **Add**.

The screenshot shows the 'Add Patients to Project' dialog box. It has a search bar at the top right. Below is a list of patient names with checkboxes: 'Example, Jeff' and 'Example, James', both of which are checked. At the bottom right are 'Cancel' and 'Add' buttons.

## Creating a Test Subject in the PureTest App

**Note:** If this option is locked, your organization's admin has disabled adding new subjects on the app. Please contact your admin if you wish to add new people for testing.

1. Tap the **Plus** icon  to access the **Add someone new** panel.
2. Select the **project** that the subject will be added to.
  - **Note:** Only projects the user has been assigned to will be displayed.
3. Add the subject's personal information. First name, last name, and project are required for all new subjects, date of birth and gender are required for shift calculations in occupational hearing testing.
  - **Note:** To promote and maintain inclusivity, a new unspecified gender has been added to the **Gender** field. Users can now select from three options: **Male**, **Female**, or **Unspecified**.
4. Add location and employment information, if applicable.
5. Add contact information and additional notes, if applicable.
6. Tap **Save** to add your new subject.



4:01 PM Wed Feb 16 31%

Cancel Add someone new Save

Fields marked with \* are required

Project\*  
Example Project

Personal

First Name\*  
Greg

Last Name\*  
Example

Date of Birth  
Required for shift calculations  
June 20, 1980

Gender  
Required for shift calculations  
Male

Location

Company


Test Prep Home Settings

**Note:** Before beginning a test workflow, ensure that the iPad is sufficiently charged. We do not recommend proceeding with a test workflow if the iPad's battery level is at 20% or lower. It is good practice to keep the iPad charging between testing sessions.


# Completing a Test Workflow

**Note:** Before beginning a test workflow, ensure that:

- The test workflow is being performed in the same quiet room that the test prep was performed in and the table and iPad are not located near any vents, fans, or appliances that emit noise (refrigerators, etc.).
- The iPad is placed **in the provided iPad stand**, with the microphone unobstructed. This will ensure that the noise monitoring functions correctly. The iPad should **not** be placed flat on the table or held in one's hands as this can cause vibrations that will trigger the noise monitoring.
- The iPad is sufficiently charged. We do not recommend proceeding with a test workflow if the iPad's battery level is at 20% or lower. It is good practice to keep the iPad charging between testing sessions.


1. On the home screen, use the search bar to find and select the subject you would like to test. If you are testing a brand-new subject, be sure to add them to your device first by tapping the **Plus** icon .

- **Note:** Before tapping the **Start Testing!** button, confirm that the test subject's demographics are correct before continuing. The panel will provide a reminder for you, highlighted in yellow.




- **Important:** If you receive either of the messages below, that means your headphone data was removed, or that your organization has blocked testing with expired headphones. This is to maintain compliance with regulatory


requirements. You will need to install newly recalibrated headphones to proceed. Contact an admin user if you do not have the new headphones.





2. If you have not yet performed the **Test Prep** activities, you will be prompted to do so.
3. Select your desired workflow. For PureTest versions 2.2 or earlier, the default workflow is the only available option. With version 2.3, it is possible to select one or more workflows depending on your needs. For example, you can set up a workflow that has a questionnaire and another that does not. Likewise, different workflows can be set up for different regulations and use cases depending on your needs.
  - **Note:** Multiple workflows are only available on PureTest version 2.3 and later. If your device is running version 2.2 or earlier, you will only have access to the default workflow.




4. Tap **Start Testing!** to begin the testing workflow. You will be asked to confirm that you are testing the right patient before continuing.




The subject will select their language and will then be brought to their personalized start page, where they will view a preview of their expected workflow.




**Note:** Tapping the **Pause** button  in the top-right corner of the screen will prompt you to enter your passcode and provide you with an overview of the workflow's current status. If you need to end the workflow for any reason, you can tap **End Workflow** to discard the current workflow and return to the home screen.

It is important to remember that the workflow cannot be resumed once exited, and exiting the workflow before the results screen will not generate a results report PDF, though questionnaire and audiogram data will be synced.





5. If the pre-test questionnaire is enabled, the subject will first perform the pre-test questionnaire.




6. If the workflow is configured with an automated test (default), the subject will be presented with a short tutorial on how to complete the test once they have completed the questionnaire. It will explain that they must drag the blue disc to either the green **sound speaker** icon if they heard a tone, or the red **no sound speaker** icon if they did not hear a tone.
  - **Note:** The tutorial is optional and can be toggled off in the web portal by an admin user.
  - **Note:** Ensure the patient has the headphones comfortably fitted onto their head, with the red cup on the right ear.

If the workflow is configured with a manual test, skip to step 10.





7. After viewing the tutorial, the subject will then begin the automated test. If the noise monitor  detects excessive background noise, it will pause the test and notify the user.

- **Important:** Do not place the iPad flat on a table. When using the provided iPad stand, do not place the iPad at an angle that could reduce the reliability of the microphone at measuring background noise. For optimal results, the iPad should be placed horizontally and evenly on the provided iPad stand when performing automated hearing tests and Room Scans.




8. During the workflow, you may notice that your subject is having difficulty completing the hearing test. If the test detects any abnormalities, or if the background noise does not subside, you will be required to enter your passcode and assist with the workflow.

- **Note:** The test can also be manually paused by tapping **Pause**  in the top-left corner of the test screen.



- If there are difficulties with the automated test, a good troubleshooting option to complete the workflow quickly is to toggle the **Assisted Mode** switch and proceed using Assisted Mode.


In Assisted Mode, you present each tone manually to the subject. You will ask the subject to face away from you and ask them to provide an indication if they heard the tone, such as raising their hand. The frequency and volume of each tone will be displayed below the noise monitor at the top of the screen.



- If the workflow is configured with a manual test, the subject will be prompted to return the iPad to the examiner after completing the questionnaire. The examiner will have to enter their PIN to continue and will then be prompted to tell the patient to put on the headphones. The test will then begin.

- **Note:** Ensure the patient has the headphones comfortably fitted onto their head with the red cup on the right ear.




11. On the manual test screen, the examiner has various controls available at the bottom of the screen as well as the display in the audiogram view above. The controls in the bottom-left of the screen are used to select the target presentation level and frequency. Once these are selected, click **Play** in the bottom right to present that tone to the subject.




12. The controls near the **Play** button allow the examiner to record the subject responses onto the audiogram.

13. Click **Complete Test** to save the current responses and proceed to the results section. Click **Quit Test** to discard all responses. If any frequencies were configured as recommended but not completed, the examiner will be reminded before exiting the test.

14. Once the test is complete, the subject will be thanked for taking the test and asked to remove the headphones. They will be prompted to return the iPad to the test administrator to proceed to the results screen.




**Note:** If frequencies were skipped, or the test was too noisy, you will be notified.



15. To proceed to the results screen, you will need to enter your passcode. You will then be able to review the test results with the subject.

- **Note for admin users:** For info on what can be configured on this page, refer to [Results](#).




Please sign below



Results can be viewed in both the examiner and subject's language if multiple languages are enabled.



An easy-to-read results summary can be found at the bottom of the audiogram, along with a symbol legend and a timestamp for when the last room scan was passed.



16. If signatures are enabled, both the subject and the test examiner will be prompted to enter their confirmation signatures after reviewing the test results. When both signatures have been entered, tap **Continue**.

Subject: Example, John

Examiner: Kramer, Andrew

By signing, you agree that you have been shown your hearing test results. If there are next steps, you have been told how to complete them.

By signing, you confirm that the test is complete and the results have been shared with the subject who was tested. If next steps are indicated, they have been communicated to the subject.

John, sign here

Examiner, sign here

Subject: Example, John

Examiner: Kramer, Andrew

By signing, you agree that you have been shown your hearing test results. If there are next steps, you have been told how to complete them.

By signing, you confirm that the test is complete and the results have been shared with the subject who was tested. If next steps are indicated, they have been communicated to the subject.

Change signature

Change signature

You're done signing! Continue to find options for sharing.

Continue

17. The final screen of the workflow will ask if the subject would like a copy of their test results. If they would like a copy, you can provide them with a physical copy by tapping on **Print** and printing their test results, or you can provide them with a digital copy by tapping on **Email**.

1:58 PM Mon Oct 19 89%

Would the subject like a copy of their results?

Print

Email

Please confirm you are finished with results

Finished

If sending the subject their results via email, you will need to confirm their personal information and email address with them. If no date of birth or email address was entered when the patient was created, you will need to enter it now. The subject's personal information will act as security information that the subject will need to access the document.

**Important:** Organizations with locked demographics enabled **will not** be able to edit the subject's information at this stage. If the subject is missing their date of birth and/or email, the option to send results via email will be grayed out, and an admin user will need to send the subject their results using the SHOEBOSX Data Management Portal. If the subject's DOB/email are filled in, the option to send via email will still be present, but the option to edit will be locked out.

Cancel      Email results      Edit

Results will be shared through a link to the subject's email.

Please confirm subject's personal information. This will act as security information which the subject will need to provide to open their results document.

Security information

First Name: Camilla  
Last name: Sanchez  
Date of birth: Tap Edit to add date of birth  
Email: Tap Edit to add email

Send now

Please confirm you are finished with the results      Finished

< Back      Security information      Save



The patient's demographics will be updated if this is edited and saved.

First name: Camilla  
Last name: Sanchez  
Date of birth\*      Fields marked with \* are mandatory  
1977 - 01 - 06  
Email\*  
c.sanchez@gmail.com


Please confirm you are finished with the results      Finished

After the information has been updated, tapping on **Send Now** will send the email to the customer's address once your data has been synced to the cloud. The subject will receive an email with a link to their results. The email will be sent by [no-reply@shoebox.md](mailto:no-reply@shoebox.md). Please inform the subject to check their spam or junk folders in case the email does not appear in their inbox.

**Note:** You can always send the patient their results from the web portal later on if the app sync fails and the email request is lost, or if the subject does not request a copy of their results during the test but changes their mind later on.



18. Tapping **Finished** will return you to the workflow overview screen. Tapping **Return Home** will complete the workflow and return you to the home screen




# Managing Test Results

SHOEBOX PureTest will automatically cache and save test results, then remove them from your device once they have been safely synced to your web portal. However, to protect the personal health information of your subjects, test results are not viewable on the iPad, only on your web portal.

Once a patient's results have been synced to your Data Management Portal, they are removed from your iPad's cache memory. After the patient's results have been synced, you can view them at any time on the **Patients** page in the Data Management Portal.


## To View Saved Results Data

- 1. Log in to your [SHOEBOX Data Management Portal](#).
- 2. Hover over **Patients** in the sidebar and click **All Patients**.
- 3. Locate the subject whose results you want to view and click on their name.



You will then be taken to that subject's individual results page, where all their previous test data can be located, along with all their other relevant information.

**Note for admins:** You can also use this page to edit the subject's demographic information and audiogram data, as well as set a next action for the subject (follow-up, update baseline, etc.).



# Exporting Results

You can export PureTest results in bulk as a ZIP folder from the patient page (all users) and the project page (admin users only).

## To Export Bulk Results Data

1. Access the **Patients** page (or **Projects** page if you want to export bulk result data from a single project only).
2. In the top-right corner, click the **Export** button.
3. Select the format you would like to export.

The screenshot shows the 'Patients' page interface. At the top left is the title 'Patients'. To its right is a search bar and an 'EXPORT' button with a dropdown arrow. Below the title is a '+ Filter Criteria' button. The main content is a table with columns: Patient Name, Company, Last Tested, and Unit. The table lists six example patients. An 'EXPORT' dropdown menu is open, showing options for ZIP and CSV formats. The ZIP options are: Individual PureTest Result Reports, Referral Letters & Summary Reports, and Individual Summary Reports. The CSV options are: Speech Discrim Results, Audiograms, Patients List, SRT Results, Hearing Conservation 3 QNR, and HHIA-S QNR.

<input type="checkbox"/> Patient Name	<input type="checkbox"/> Company	<input type="checkbox"/> Last Tested	<input type="checkbox"/>
<input type="checkbox"/> <a href="#">Example, Roger</a>	Example Company	2020-11-11	Ur
<input type="checkbox"/> <a href="#">Example, Jeff</a>	Example Company	2020-11-18	Ur
<input type="checkbox"/> <a href="#">Example, John</a>	Example Company	2021-04-20	Ur
<input type="checkbox"/> <a href="#">Example, Bob</a>	Example Company	2021-06-08	Unit 1
<input type="checkbox"/> <a href="#">Example, Richard</a>	Example Company	2021-06-08	Unit 1
<input type="checkbox"/> <a href="#">Example, Joe</a>	Example Company	2021-06-18	Unit 1

The screenshot shows the 'Projects' page interface. At the top left is the title 'Example Project'. To its right is an 'EXPORT' button with a dropdown arrow. Below the title is a 'Project Details' section with a table of project information. An 'EXPORT' dropdown menu is open, showing options for ZIP and CSV formats. The ZIP options are: Individual PureTest Result Reports, Referral Letters & Summary Reports, and Individual Summary Reports. The CSV options are: Audiogram PDFs, Audiograms, Speech Discrim Results, Patients List, SRT Results, Hearing Conservation 3 QNR, and HHIA-S QNR.

Project Details	
Start date	Oct 16, 2020
End date	Oct 16, 2028
Status	Open
Available for PureTest offline testing	Yes
Description	Example Description


# Test Prep Reports

Test Prep Reports allow admin users to audit their own test examiners to ensure that the mandatory test prep activities are being completed, and to monitor how often each PureTest unit in their organization passes or fails these checks.


This feature also allows admin users to quickly and easily provide a full summary of all test prep activity completion in CSV format if records of these checks are requested by an auditor.

## To Generate a Test Prep Report


1. Log in to your [SHOEBOX Data Management Portal](#).
2. Click on **Reports** in the sidebar.
3. Select the test prep activity you would like to generate a report for under the **Test Prep Reports (PureTest)** heading. For this example, we will use the **Room Scan Report**.




4. Use the dropdown menus to configure the date range for the report, and whether you want the report to generate total passed checks, failed checks, or both.



5. Click on **Apply** to generate the report.



6. Click the **Export** button to generate a CSV with detailed information from the report. This CSV export contains all the necessary information required for you to keep on file in case of an audit




You can then repeat this process for the Headphones Check Report and Biological Verification report.

# Setting Device Security Options

Using your SHOEBOX Data Management Portal, you can configure the security settings for all devices within your organization.

## To Access Your Device Security Options

1. Log in to your [SHOEBOX Data Management Portal](#).
2. Hover over the **Settings** tab in the sidebar.
3. Click **Security**.




## Lock Screen

This section allows you to control how long it takes for the PureTest app's lock screen to appear when the iPad is left untouched. This is done to protect patient health information by locking the screen when the app is not in use. The lock screen will not interfere with a test subject's workflow.

You can also choose to disable the lock screen entirely by turning off the toggle switch, though this is not recommended.

After adjusting your lock screen settings, click **Save** to save your changes.

**Note:** The default lock screen setting is 2 minutes. To use this default, simply turn on the toggle switch and click **Save** to save your changes.



## Block Testing with Expired Headphones

This feature allows you to prevent test examiners from testing with expired headphones. Enabling this feature ensures accurate test results and compliance with any regulatory requirements you may need to adhere to.

After enabling this feature, be sure to click **Save** in the top-right corner of the **Security** page to save your changes.

### Block Testing with Expired Headphones




*Headphones cannot be used for testing if their calibration status is expired.*

# Configuring Language Settings


Using your SHOEBOX Data Management Portal, you can configure the language settings for all devices within your organization. Only English is configured by default. If your test examiners and test subjects are fluent in English, simply leave the default settings as they are. This shortens the testing workflow by removing a step where test subjects are asked to choose a language for the test. However, if you want test subjects to choose from multiple languages, or you would like to change the default language to a language other than English, you need to change the default settings.

## To Change the Default Language Settings

1. Log in to your [SHOEBOX Data Management Portal](#).
2. Hover over the **Settings** tab in the sidebar.
3. Click **Language**.




4. On the **Language** page, click **Edit**.
5. Under **Test Subject Languages**, select the languages that you would like to be enabled or disabled.




6. Click **Save** to save your changes.
  - **Note:** If your organization uses custom text for the subject/examiner signatures messages, you will receive this notification recommending that you review the signature text for the other languages you have enabled. You may be required to provide a custom translation for the additional languages you have enabled if

you want them to exactly match the custom text you have provided for your default language.




7. Once saved, the app will be configured to prompt the patient at the start of the testing workflow to select the language they would like to use.



# App Settings

The **Settings** screen contains all the details on your app's hardware, test settings, offline projects, and other miscellaneous information.



## Hardware

The Hardware section is where you will go to install, change, or update your headphones and microphone. It will also provide you with a notification if your hardware is close to its expiration date.


**Note:** If you are using an iPad that does not have a 3.5mm headphone jack, you must use an Apple USB-C to 3.5mm Headphone Jack Adapter to connect the headphones. Currently, PureTest only supports Apple's adapter.


### To Change or Update Your Headphones:

1. Tap on **Headphones & Microphone**.

The **Headphones** panel will display the name of the headphones that are currently installed on your iPad. If none are installed, the panel will be blank and indicate that you


have not installed any headphones yet.



2. Tap the **Plus** icon  to open the **Change Headphones** panel.

You will be able to select new headphones from the list of available headphones, or by using the search bar to search for your headphones by their 12-digit ID number printed on the label attached to the wire of your headphones.


Alternatively, you can tap the **Barcode** icon  to scan the QR code attached to the label on the wire of your headphones.



### To Change or Update Your Microphone:


1. Tap **Headphones & Microphone**.
2. The **Microphone** panel will display the name of the microphone that is currently installed on your iPad. If none are installed, the panel will be blank and indicate that you

have not installed a microphone yet.



3. Tap the **Plus** icon  to open the **Change Microphone** panel.


You will be able to select a new microphone from the list of available microphones. Alternatively, you can use the search bar to search for your microphone by the 6-digit ID number printed on the side of your microphone.



## Removing a Microphone

If, for any reason, you need to remove your microphone, simply tap **Remove** on the **Microphone** panel.

**Important:** A microphone is required for noise monitoring when testing under OSHA regulations. After removing a microphone, a message will appear in the Microphone panel to remind you of this requirement.



## Testing


This section provides details on how your app has been configured by your organization's admin user, and allows you to download any updates to your settings that may have been made while you have been logged in.

**Note:** Admin users are responsible for using the SHOEBOS Data Management Portal to configure device settings, along with providing access to projects/making them available offline.

### Download Device Settings

This screen shows you which device settings have been configured by your organization's admin user, and the date/time of when these settings were last downloaded to your device.

To update them immediately, tap **Download**.




**Note:** Device settings are automatically downloaded **every hour** while you are connected to the internet. However, if you know that the admin user has made recent changes to your device settings, it is recommended that you download the new settings immediately to stay up to date before you test your next subject.

## Testing Offline

This screen shows you which projects have been marked for offline testing.

When your iPad is online, tapping **Download** will sync the latest data from your web portal. This will enable you to perform offline testing on any subjects in the downloaded project(s).




**Note:** Data will be automatically synced to the device **every four hours** if you are connected to the internet. However, if you know that you have been added to a new project, and/or you may be going offline soon, be sure to manually sync your data by tapping **Download**.

## Workflow


This screen displays the workflow settings currently configured for your organization. If workflow settings have been changed, they will be reflected on this screen after downloading them using the **Download Device Settings** screen.

**Note:** This screen is read only, and workflow settings cannot be changed on the app. Contact an admin user to request changes.



## About

This section contains miscellaneous app information including the latest feature updates, your app's current version, legal information, and developer credits.



## Feature updates

Contains information about the latest feature updates for PureTest.

**Note:** Feature updates will be stored for a limited time. If you have any questions about updates that are no longer listed, contact our [support team](#).

**Version**

Contains information about the current version of your PureTest app.

**Note:** Also includes your app's installation ID, which is used for troubleshooting application issues with SHOEBOX customer support.

**Terms of Service**

Allows you to view the current version of the SHOEBOX terms of service.

**Privacy Policy**

Allows you to view the current version of the SHOEBOX privacy policy.

**Acknowledgements**

Provides credit for third-party software used by SHOEBOX in development.

# Troubleshooting

If you experience problems while using PureTest, please review this section and attempt to troubleshoot the issue by following the steps provided. If your issue is not described here or persists after attempting to resolve it, please [contact support](#).

- [App is Running Slowly](#)
- [Blocked for Testing Error](#)
- [Checking and Updating Your PureTest App](#)
- [Connecting Transducers to a 10th Gen iPad](#)
- [Crackling Noise from Transducer](#)

## App is Running Slowly

If you notice that the performance of your app is subpar, or that the app is running slower than usual, this is most likely due to large amounts of data stored locally on the device.

### Some troubleshooting options:




1. If you have been performing offline testing, and it has been a while since you last re-connected your iPad to the internet, you may need to do so. PureTest will save and cache test results on the iPad until they can be synced to your organization's cloud storage.  
However, if your iPad is constantly offline, the app will never get the chance to perform a sync, causing a large volume of data to be cached. This can impact performance. Reconnecting your iPad to the internet will perform a sync with your cloud storage right away, and will transfer all of the cached test data to your organization's secure web portal, storing it there rather than on the app.
2. If you have not been performing offline testing, or if you perform offline testing but regularly reconnect your iPad, the issue may be coming from outside of the PureTest app, on the iPad itself. A couple key things to check:
  - How much storage space is left on the iPad? Are there other apps, files, or photos taking up space?
  - Has the iPad been dropped, hit, or otherwise damaged?
  - Is your version of the PureTest app up to date? You may need to download an update from the app store.
  - Is the system software up to date? You may need to go to your iPad's Settings app and perform a software update.

# Blocked for Testing Error

You received an error message stating that your iPad is temporarily blocked for testing.

## Why did I receive this message?

There are three reasons why you may receive this error message. See the table below for descriptions of each reason.

Error Message	Reason
 <b>iPad temporarily blocked for testing</b> <i>This iPad is temporarily blocked for testing because the current iPadOS version is not compatible with PureTest. To unblock, update iPadOS to the latest version on your app store.</i>	The current iPadOS version is not compatible with PureTest.
 <b>iPad temporarily blocked for testing.</b> <i>This iPad is temporarily blocked for testing because the current PureTest version is not compatible with the current iPadOS. To unblock, update PureTest to the latest version on your app store.</i>	The current PureTest version is not compatible with the current iPadOS version.
 <b>iPad temporarily blocked for testing.</b> <i>Your iPad is currently blocked for testing as both PureTest and iPadOS versions are not compatible. To unblock, update both versions the latest one in your app store.</i>	The current versions of both PureTest and iPadOS are not compatible.

For more information about which versions of PureTest and iPadOS are compatible, see [System Requirements & Supported Browsers](#).

## How do I resolve this issue?


Depending on the error message you receive, you must update your version of PureTest, iPadOS, or both to resolve the issue.


- To learn how to update your version of the PureTest app, see [Checking and Updating Your PureTest App](#).
- To learn how to update your version of iPadOS, see [Update your iPhone or iPad](#) on the Apple Support website.

## What if my versions are compatible and I received this error?


If you updated your installed PureTest app to the latest version, but did not start the application while connected to the internet, you may receive the blocked for testing error when attempting to test offline. To remove the block, you must connect to the internet and update your PureTest device settings.

## To update your PureTest device settings


1. Make sure your iPad is connected to the internet.
2. Open the PureTest app and log in.
3. In the top-left corner of the PureTest home screen, tap the **Internet Connection** icon  to check whether the PureTest app is connected to the internet. If not connected, tap **Re-check**.



4. In the bottom-right corner of the PureTest home screen, tap **Settings**.
5. Under **Testing**, tap **Download Device Settings**.



6. On the **Download Device Settings** screen, tap **Download**. Your device settings will be updated to the latest version.



# Checking and Updating Your PureTest App

Using an out-of-date app version has the potential to negatively impact the functionality of your SHOEBOS software. To keep your SHOEBOS app working correctly, it is recommended that you always keep your app up to date. Similarly, if you notice your app is encountering frequent errors, updating your app may solve the problem.

**Note:** Customers using SHOEBOS Mobile Device Management services should [contact support](#) for any questions or concerns regarding updates.

## Checking the Version of Your PureTest App

There are two ways to check what version of the PureTest App you are running:

1. From within the SHOEBOS PureTest app:
  - a. Tap Settings in the bottom-right corner of the **Home** screen.
  - b. The About section will have the current version listed. Tap Version for more detailed information.
2. From the Apple App Store, search for SHOEBOS PureTest and check the button next to the app.
  - a. If the button says Update, then you are not running the current version.
  - b. If the button says **Open**, then your system is up to date.

## Updating the Version of Your PureTest App

First, you need to verify which version of iOS/iPadOS (Operating System) is running on your iPad.

1. From your iPad's home screen, tap the **Settings** icon.
2. On the **Settings** screen, tap **General**, then tap **About**.
3. Find **Software Version** and take note of the iOS/iPadOS version number that is currently installed on your iPad.

If your iOS/iPadOS version is not compatible, or is out of date, you will need to update to the latest version of iOS/iPadOS before you can proceed. To learn how to update your version of iPadOS, see [Update your iPhone or iPad](#) on the Apple Support website.

Once you've ensured that your iOS/iPadOS version is compatible with SHOEBOS, you can update your app.

1. Search the App Store for SHOEBOS.
2. Tap on the version(s) of the app you are using (in this case, SHOEBOS PureTest).
  - If the button next to the app says **Update**, then you are not running the current version. Tap on **Update** to begin downloading the latest version of the app.
  - If the button next to the app says **Open**, then your system has the most current version of the app installed.

**Note:** You can always check the version of your app by going to the **Settings** page in the SHOEBOS app and tapping **About**.


# Connecting Transducers to a 10<sup>th</sup> Gen iPad

This issue occurs if you plug in a pair of transducers connected by an Apple DAC and tap **Other Device** in the confirmation popup. Selecting this option may result in the transducers not performing as expected. To correct the issue, you will need to follow the steps below depending on whether the transducers are still plugged in to the iPad.


## Device is still plugged in

If the device is still plugged in to the iPad, you can set the device type in the iPad's sound settings. To do this, follow the instructions below:


1. On your iPad, go to **Settings > Sounds > Headphone Safety**.



2. On the **Headphone Safety** screen, tap **USB Audio Accessories**.




3. On the **USB Audio Accessories** screen, set the **Connected to Headphones** toggle to match your device:
  - If the device is a pair of transducers, set the toggle on.
  - If the device is a microphone, set the toggle off.




## Device is no longer plugged in

If the device is no longer plugged in to the iPad, you will need to remove all devices assigned to the iPad's memory. To do this, follow the instructions below:


1. On your iPad, go to **Settings > Sounds > Headphone Safety**.




2. On the **Headphone Safety** screen, tap **USB Audio Accessories**.



3. On the **USB Audio Accessories** screen, tap **Forget All USB Audio Accessories**.



4. In the confirmation popup, tap **Forget All USB Audio Accessories**. This removes all assigned audio devices from the iPad's memory.



5. You can now plug your devices in to the iPad again.

## Crackling Noise from Transducer

1. Check the transducer for physical damage, such as cuts on the cable or audio jack.
2. Unplug the transducer, plug it back in, and gently wiggle the cord into the audio jack before testing again.
3. Try plugging the transducer into another device, such as a cell phone, and check if the crackling occurs.
4. Try plugging a different transducer into the iPad (such as a pair of earbuds) and check to see if the crackling occurs while using those (this could indicate an issue with the iPad's audio jack).

# Reporting a Problem


If you encounter any issues or errors while using PureTest, we highly recommend taking advantage of the "Report a Problem" feature located in the Account panel.

Using this feature will automatically contact a customer support representative to assist with your issue, while also providing logging data for the SHOEBOX development team in case your issue requires some bug fixing or investigation on our end.


**Note:** This feature is intended for troubleshooting only. If you have questions about how to use PureTest, please refer to our help center. You will find a link to the Help Center above the Report a Problem feature in the Account panel.

## To Report a Problem


1. Check to make sure that your device is connected to the internet, and that PureTest is online. The report a problem feature is only available while online.
2. Tap on the username in the top-right corner of the home screen to open the **Account Panel**.
3. Tap on **Report a problem**.



4. Your email should automatically be entered into the email field, but you can enter a different email if necessary.




5. Use the dropdown menu to select the type of problem you have encountered. Select **Other...** if your problem is not on the list.
6. Tap on **Next**.



7. Provide details about your issue using the text field provided, so that the support team can prepare to assist you. If possible, we highly recommend describing the steps you took before you encountered the problem, so our support team can reproduce the issue in the software.

**Important:** For privacy protection, please do not include any patient names or test result information.


8. Tap on **Submit**.



Your problem will then be logged and will automatically open a support ticket.

A support representative will then contact you within the next business day to help resolve the issue.

If your problem needs immediate attention, please call our support team at 1-877-349-9934.



# SHOEBOX Audiometry Specifications

Audiometer Type	Type 4
Permissible environmental conditions	See iPad technical specifications available on <a href="http://www.apple.com">www.apple.com</a>
Permissible power supply variations	SHOEBOX Audiometry will only function when iPad is on battery power
Measures to minimise unwanted sound radiation	Touch screen interface, no sound is emitted
Identification of the transducers and their reference equivalent threshold levels	See transducer calibration certificate for information on transducer calibration
Ear simulator used for calibration	As per ANSI/ASA S3.6-2018
Static force provided	As per ANSI/ASA S3.6-2018
Warm up time	No warm up time required. See iPad technical specifications on <a href="http://www.apple.com">www.apple.com</a>
Sensitivities and nominal impedances of all input facilities.	No input facilities
Available voltage and nominal impedance of all output facilities	Output is CTIA/AHJ 3.5mm audio output of Apple iPad. See iPad technical specifications on <a href="http://www.apple.com">www.apple.com</a>
Pin assignment of all external plug connections	Output is CTIA/AHJ 3.5mm audio output of Apple iPad. See iPad technical specifications on <a href="http://www.apple.com">www.apple.com</a>
Mode of operation	Pure tone forced choice modality utilizing a Modified Hughson Westlake algorithm
Rate of change of sound pressure level for automatic-recording audiometers	n/a
Rate of change of frequency	n/a, not a Bekesy algorithm
Frequency of modulating signal	Warble tone frequency = 5Hz
Modulation waveform	Sine
Modulation range	6%

Sound attenuation characteristics of the earphones as measured in accordance with ISO 4869-1	For transducer sound attenuation characteristics, refer to transducer manufacturer documentation
Maximum hearing level settings provided at each test frequency including limitations in use due to harmonic distortion	SHOEBOX Audiometry PureTest = 90 dB HL No limitations due to harmonic distortion
Information about the time window for subject's response for automated test procedures	Pure tone forced choice modality does not involve subject response timing
Type of battery, means of checking the battery and method of replacement, expected battery life time	See iPad technical specifications available on <a href="http://www.apple.com">www.apple.com</a>
Maintenance and calibration procedures and schedules	Headphones must be calibrated annually per ANSI/ASA S3.6-2018 Calibration procedure verifies correct operation and identifies required maintenance
EMC warning	See iPad technical specifications available on <a href="http://www.apple.com">www.apple.com</a>
Type of auditory signal	Pure tones, warble tones, pulsed tones, masking tones, speech recordings
Type of transducers and their headband force	Various ANSI/ASA S3.6-2018 compliant transducers are available, refer to transducer manufacturer's documentation
Sound field system	None, not supported
Type of ear simulator or mechanical coupler used for calibration and the method of coupling the transducer	See calibration certificate for each transducer
Type of calibration	Calibrated in dBHL for known RETSPLs for each transducer according to ANSI/ASA S3.6-2018
Acoustic or Vibratory Signal Level for a given setting of the output-level control	Control interface displays dBHL based on RETSPL of the currently selected transducer. This is mapped to output voltage level appropriately
Polarity of resulting acoustic or vibratory signal	Left and right channel outputs produce in-phase sounds and their polarity is set during calibration
Repetition rate	Modified Hughson Westlake algorithm is used
Duration of the initial sound pressure or vibratory force wave of a click and/or duration and rise/fall times of acoustic or vibratory tone-bursts	As specified in ANSI/ASA S3.6-2018
Subjective relationship between test signals and reference signals	RETSPLs from the manufacturers are used to enable normalized hearing level outputs
Environmental conditions at time of calibration	As specified in ANSI/ASA S3.6-2018 and as recorded in annual calibration equipment certification
Storage	Store iPad and headphone accessories according to accessory manufacturer recommendations

# Appendix A: Standard PureTest Kit Inclusions

Medical Device Hardware	<ul style="list-style-type: none"><li>● DD450 Headphones (please see IFU supplied by Manufacturer)</li></ul>
Non-Medical Device Hardware	<ul style="list-style-type: none"><li>● Carrying Case</li><li>● PureTest Case Label</li><li>● iPad Aluminum Stand</li><li>● External Microphone (Optional, for Noise Monitoring only)</li><li>● iPad (Optional)</li><li>● Digital Analog Converter (Optional, for iPad 10+)</li></ul>
Documentation	<ul style="list-style-type: none"><li>● For <b>PureTest Essentials</b>, EXT-57 SHOEBOX PureTest Essentials Card is included.</li><li>● For <b>PureTest Shift</b>, EXT-58 SHOEBOX PureTest Shift Card is included.</li></ul>